

PRODUCT SPECIFICATION SHEET

Water

USP Grade

Test	Specification	Typical Results
Appearance	Clear colorless liquid	Pass
pH ¹	5.0 – 7.0	6.0
Resistivity ¹	≥10 megohm-cm	18 megohm-cm
Total Organic Carbon ¹	<0.50 mg C/L	<0.50 mg C/L
Water Conductivity ¹	NMT 1.3 µS/cm at 25C	0.05 µS/cm
Substances Reducing Permanganate	To Pass Test	Pass
Microbial Limits	≤100 cfu/mL	≤10 cfu/mL

¹ Meets stated value at the time of manufacture.

Appropriate/legal use of this product is the responsibility of the user.