

PRODUCT SPECIFICATION SHEET

Water

Meets LC/MS Monographs

Test	Specification	Typical Results
Residue after Evaporation	Max. 1 mg/L	< 1 mg/L
Ultraviolet Absorbance Test	To Pass Test	Pass
Refractive Index at 20°C	1.3325 – 1.3335	Pass
LC gradient elution suitability at 254nm	To Pass Test	Pass
Fluorescence at emission maximum for Impurities (as Quinine Sulfate)	Max. 0.30 ppb	< 0.30 ppb
LC-MS Positive Mode (as Reserpine)	Max. 50 ppb	< 50 ppb
LC-MS Negative Mode (as Nitrophenol)	Max. 50 ppb	< 50 ppb
Trace Metals	To Pass Test	Pass

Appropriate/legal use of this product is the responsibility of the user